





EU Declaration of Conformity

5550 Vibration Switch

Manufacturer's Name	Metrix Instrument Co., L.P.
Manufacturer's Address	8824 Fallbrook Drive, Houston, TX 77064 USA 281.940.1802 www.metrixvibration.com info@metrixvibration.com
Declares that the product	Model 5550 Vibration Switch (This declaration of conformity is issued under the sole responsibility of Metrix Instrument)
Directive 2014/34/EU ATEX	Provision of the Directive fulfilled by the Equipment <ul style="list-style-type: none">•  II 2 GD• Ex d IIB +H₂ T6 Gb Ta -40°C to +70°C• Ex tb IIIC T85°C Db Ta -40°C to +70°C IP66
Conforms to the following standards	ATEX <ul style="list-style-type: none">• EN 60079-0:2012
Other Specification Standards	<ul style="list-style-type: none">• EN 60079-1:2007- This standard has been compared with EN 60079-1:2014 (currently harmonized) and no significant changes have occurred which are applicable to this equipment
EU-Type Examination	Certificate <ul style="list-style-type: none">• Baseefa10ATEX0098• SGS Baseefa Limited Rockhead Business Park, Staden Lane Buxton, Derbyshire SK17 9RZ UNITED KINGDOM Notified Body Number: 1180
Notified Body for production	SGS Baseefa Limited (Notified Body Number: 1180) Buxton, UK
Technical File Location	The Technical File required by this Directive is maintained by the manufacturer listed above and the address listed above.
Approved By	 Lance Truong Quality Manager, Metrix Instrument Co. Houston, Texas 77064 August 20, 2017