

제 2014-BO-0130 호



# 안 전 인 증 서

Metrix Instrument Co.

8824 Fallbrook Brive, Huston, TX, USA 77064

위 사업장에서 제조하는 아래의 품목이 산업안전보건법 제34조 및 같은 법 시행규칙 제58조의4제4항에 따른 안전인증 심사 결과 안전·보건기준에 적합하므로 안전인증표시의 사용을 인증합니다.

## 품 목

ViBration Mechanical Switch

## 형식·모델/용량·등급/인증번호

형식·모델

5550-4\*\*-\*\*\*

인증번호

14-AV4BO-0130

Ex d IIB+H2 T6 IP65

용량·등급 Micro switches: 480 Vac, 15 A, Coil: 230 Vac, 1 A / 24 Vdc, 3 A,  $-40\text{ }^{\circ}\text{C} \leq T_a \leq +70\text{ }^{\circ}\text{C}$ , 동일형식일람표 참조

## 인 증 기 준

방호장치 의무안전인증 고시(고용노동부고시 제2010-36호)

## 인 증 조 건

Metrix Instrument Co., 8824 Fallbrook Brive, Huston, TX, USA 77064에서 생산되는 제품에 한함.

2014년 02월 27일

# 한국산업안전보건공단 이사장



# 동 일 형 식 일 람 표

(1/ 총1)

사업장명	Metrix Instrument Co.		개정일자 및 번호	-	인증번호	14-AV4BO-0130	
일련 번호	구조 · 등급	형식 · 모델	인정항목 및 내역				비고
1	Ex d IIB+H2 T6 IP65	5550-4XX-XXX	Relay Contacts  1 = SPDT, Silver, 15A 2 = DPDT, Silver, 15A 3 = SPDT, Gold, 1A 4 = DPDT, Gold, 1A	Reset / Startup Delay Coil Holding Strength Resett / Startup Delay Coil Voltage  0=None 1=115Vac 2=230Vac 3=24Vdc 4=115Vdc	Wiring Entry / Mounting Hole Pattern  1= 3/4"NPT / Metrix 5173 or 5175 2= 3/4"NPT / Metrix 5097, VS-2 EX or 366 3= 3/4"NPT / Metrix 5078 or 365 4= M20x1.5 / Metrix 5097, VS-2EX, or 366 6= M20x1.5 / Metrix 5173 or 5175 7= 3/4"NPT Metrix or PMC-BETA 440 8= M20x1.5 / Metrix 5078 or 365	Environmental Rating  0=NEMA 4 / IP66 1=NEMA4X / IP66	(2014.02.27)